



## Biomaterial modification of urinary catheters with antimicrobials to give long-term broadspectrum antibiofilm activity

Leanne E. Fisher<sup>a,1</sup>, Andrew L. Hook<sup>b</sup>, Waheed Ashraf<sup>a</sup>, Anfal Yousef<sup>a</sup>, David A. Barrett<sup>c</sup>, David J. Scurr<sup>b</sup>, Xinyong Chen<sup>b</sup>, Emily F. Smith<sup>d</sup>, Michael Fay<sup>d</sup>, Christopher D.J. Parmenter<sup>d</sup>, Richard Parkinson<sup>e</sup>, Roger Bayston<sup>a,\*</sup>

<sup>a</sup> Biomaterials-Related Infection Group, School of Medicine, Nottingham University Hospitals, Queen's Medical Centre, Nottingham NG7 2UH, UK

<sup>b</sup> Laboratory of Biophysics and Surface Analysis, School of Pharmacy, University of Nottingham, Nottingham NG7 2RD, UK

<sup>c</sup> Centre for Analytical Bioscience, School of Pharmacy, University of Nottingham, Nottingham NG7 2RD, UK

<sup>d</sup> Nottingham Nanotechnology & Nanoscience Centre, University of Nottingham, Nottingham NG7 2RD, UK

<sup>e</sup> Nottingham Urology Centre, Nottingham University Hospitals NHS Trust, Nottingham NG5 1PB, UK

### ARTICLE INFO

#### Article history:

Received 22 September 2014

Accepted 28 January 2015

Available online 30 January 2015

#### Keywords:

Antimicrobial

Bladder

Catheter infection

Drug release

Silicone

Urinary tract

### ABSTRACT

Catheter-associated urinary tract infection (CAUTI) is the commonest hospital-acquired infection, accounting for over 100,000 hospital admissions within the USA annually. Biomaterials and processes intended to reduce the risk of bacterial colonization of the catheters for long-term users have not been successful, mainly because of the need for long duration of activity in flow conditions. Here we report the results of impregnation of urinary catheters with a combination of rifampicin, sparfloxacin and triclosan. In flow experiments, the antimicrobial catheters were able to prevent colonization by common uropathogens *Proteus mirabilis*, *Staphylococcus aureus* and *Escherichia coli* for 7 to 12 weeks in vitro compared with 1–3 days for other, commercially available antimicrobial catheters currently used clinically. Resistance development was minimized by careful choice of antimicrobial combinations. Drug release profiles and distribution in the polymer, and surface analysis were also carried out and the process had no deleterious effect on the mechanical performance of the catheter or its balloon. The antimicrobial catheter therefore offers for the first time a means of reducing infection and its complications in long-term urinary catheter users.

© 2015 The Authors. Published by Elsevier B.V. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

### 1. Introduction

Urethral catheters are used to drain urine from the bladder (Fig. 1). Bladder catheterization is commonly required, either for short-term bladder drainage or for long-term management of bladder dysfunction, and at least 25% of hospital patients will have a bladder catheter placed at some point in their stay. Short-term catheters are used for the temporary relief of reversible bladder voiding difficulties, for urine output monitoring or after lower urinary tract surgery and are typically used


for between 1 and 14 days. Long-term catheterization may be used to manage intractable urinary problems such as chronic urinary retention or incontinence not treatable by other means, and here the aim is to keep the catheter functioning and infection-free for as long as possible but while medical opinion varies, and despite careful hygiene in handling they usually need to be changed every 4–8 weeks [1]. Catheter-associated urinary tract infection (CAUTI) is the commonest hospital-acquired infection, accounting for 40% of all nosocomial infections and over 100,000 admissions to hospital within the USA annually [2]. CAUTI rates have continued to rise in almost every care unit type [3].

The most common CAUTI pathogen is *Escherichia coli*, followed by *Proteus mirabilis* [4]. Others, such as *Klebsiella pneumoniae*, *Enterobacter* spp, enterococci and staphylococci are important but less common, and *Pseudomonas aeruginosa* and *Candida* spp may be seen in longer-catheterized patients, particularly after repeated courses of antibiotics. Increasingly, strains of *E. coli* and *K. pneumoniae* produce extended spectrum beta-lactamases (ESBL), making them insusceptible to even the newer cephalosporin antibiotics and presenting added therapeutic challenges. As in other devices, CAUTI pathogens are able to attach to the catheter material and to develop biofilms. CAUTI due to *P. mirabilis* is

\* Corresponding author at: Biomaterials-Related Infection Group, School of Medicine, C Floor West Block, Nottingham University Hospitals, Queen's Medical Centre, Nottingham NG7 2UH, UK.

E-mail addresses: [le.fisher@btinternet.com](mailto:le.fisher@btinternet.com) (L.E. Fisher), [Andrew.hook@nottingham.ac.uk](mailto:Andrew.hook@nottingham.ac.uk) (A.L. Hook), [waheed.ashraf@nottingham.ac.uk](mailto:waheed.ashraf@nottingham.ac.uk) (W. Ashraf), [anfalyousef@gmail.com](mailto:anfalyousef@gmail.com) (A. Yousef), [david.barrett@nottingham.ac.uk](mailto:david.barrett@nottingham.ac.uk) (D.A. Barrett), [david.scurr@nottingham.ac.uk](mailto:david.scurr@nottingham.ac.uk) (D.J. Scurr), [x.chen@nottingham.ac.uk](mailto:x.chen@nottingham.ac.uk) (X. Chen), [emily.smith@nottingham.ac.uk](mailto:emily.smith@nottingham.ac.uk) (E.F. Smith), [Michael.fay@nottingham.ac.uk](mailto:Michael.fay@nottingham.ac.uk) (M. Fay), [Christopher.parmenter@nottingham.ac.uk](mailto:Christopher.parmenter@nottingham.ac.uk) (C.D.J. Parmenter), [Richard.parkinson@nuh.nhs.uk](mailto:Richard.parkinson@nuh.nhs.uk) (R. Parkinson), [roger.bayston@nottingham.ac.uk](mailto:roger.bayston@nottingham.ac.uk) (R. Bayston).

<sup>1</sup> These authors contributed equally to this work.



**Fig. 1.** Schematic of a catheterized urinary bladder, showing the Foley catheter in place. The catheter is inserted into the bladder through the urethra, and the balloon is then inflated with water using a syringe attached to a secondary channel in the catheter. The function of the balloon is to prevent the catheter from slipping out of the bladder. Urine enters the bladder from the kidneys via the ureters, and drains into the eye-holes in the catheter and then into a collecting bag.

especially important due to associated biomineralization [4,5] that can block the catheter lumen, causing obstruction and risking kidney infection and septicemia.

At least two approaches for prevention of biomaterial infection have been used. One involves modification of the biomaterial surface to reduce bacterial attachment, a pre-requisite event in biofilm development. This is usually aimed at making the biomaterial surface hydrophilic [6,7] but a class of weakly amphiphilic polymers that resist bacterial attachment have also been identified [8,9]. The second approach has been to attach active biocides such as antibiotics to biomaterial surfaces, or to impregnate them into the biomaterial itself. One example of the use of surface biocides is silver-processing using various techniques [10]. However, these have been disappointing in clinical use [11] and few “anti-biofilm” urinary catheters have reached the market. Urinary catheters containing nitrofurazone have been evaluated in a large randomized controlled clinical trial alongside silver-processed and plain catheters, and neither of these “antimicrobial” commercially available catheters showed a clinically significant reduction of infection [12]. International guidelines now state that the evidence is not sufficient to support their use in short-term (<30 days) or long-term (>30 days) users [13] and nitrofurazone catheters are now no longer available. In general, antimicrobial coatings either are depleted rapidly by urine flow, or become obliterated by a host protein conditioning film. We have previously reported an antimicrobial neurosurgical catheter produced by an impregnation process that has a long duration of activity confirmed by thousands of successful implants [14–17] and the process has been adapted for use in dialysis [18]. Our previous research on dialysis catheters, using a different antimicrobial combination, concentrated on *in vitro* assessment of antimicrobial activity and included investigation of potential for inflammatory reaction in the extremely sensitive peritoneal cavity. Here we report the evaluation of duration of activity of a different antimicrobial combination, drug concentrations and their distribution in the catheter material as well as their release characteristics, effect of processing on mechanical properties, particularly of the important retention balloon, and surface analysis on a novel antimicrobial catheter intended for long-term urinary drainage, with enhanced antimicrobial spectrum and duration of activity. The antimicrobials chosen, rifampicin, sparflaxacin and triclosan, were chosen for their spectrum of activity against CAUTI pathogens (rifampicin

and triclosan against staphylococci, sparflaxacin and triclosan against *E. coli*, *K. pneumoniae* and *P. mirabilis*). The choice was also governed by their physicochemical characteristics: solubility in chloroform and ability to diffuse through the crosslinked silicone matrix.

## 2. Materials and methods

### 2.1. Impregnation process

The impregnation process (Fig. 2) was carried out as described previously [14]. The antimicrobials, rifampicin (Sigma-Aldrich, Poole, UK), triclosan (Ciba Speciality Chemicals, Macclesfield, UK) and sparflaxacin (Sigma-Aldrich) were chosen for their activity against the target bacteria (CAUTI pathogens) and their chemical compatibility with the impregnation process. Briefly, the antimicrobials were dissolved together in chloroform (Fisher Scientific, Loughborough, UK) to give concentrations w/v of 0.2% rifampicin, 1% triclosan and 1% sparflaxacin. Foley catheters/segments (Coloplast, Peterborough, UK) or silicone test discs 1 mm × 6 mm (Goodfellow, Cambridge, UK) were immersed in the solution for 1 h, during which the silicone swelled to approximately twice its volume. The catheters and discs were then removed and rinsed in absolute ethanol (Fisher Scientific) to remove residual solvent and drug, and allowed to dry overnight at room temperature in a current of air. During evaporation of the solvent the catheters returned to their previous dimensions (Fig. 2), leaving the antimicrobials distributed evenly throughout the silicone matrix. Separately, a series of silicone test discs was produced containing the above antimicrobial concentrations as single agents. The segments and discs were then packaged and sterilized by autoclaving at 121 °C for 15 min. The sterilization process had no significant effect on the antimicrobial activity (data not shown).


### 2.2. Assay of drug content and drug release profiles

Total drug content was determined by extracting catheter segments in chloroform which was then evaporated at room temperature. Drug residues were re-dissolved in acetonitrile and HPLC analysis was performed (Agilent 1090 HPLC, Agilent Technologies, Berkshire, UK) (Supplementary Method 1). The total concentration of each drug extracted from the catheter segments was calculated using peak areas from calibration curves and the total drug content per catheter was calculated. All experiments were carried out in triplicate. Calibration curves showed good linearity, with correlation coefficients ( $R^2$ ) for rifampicin, triclosan and sparflaxacin of 0.9961, 0.9955 and 0.9997 respectively. To establish drug release concentrations, antimicrobial catheter segments were placed into HPLC grade water (pH 7) (Fisher Scientific) in a 37 °C incubator with constant agitation. Segments were transferred daily or every 2–4 days into fresh water over a period of 28 days. After concentration by liquid–liquid extraction using chloroform, eluates were evaporated and drug residues were re-dissolved and analyzed by HPLC. Drug release concentrations were determined from calibration curves. All tests were carried out in triplicate. Calibration curve correlation coefficients ( $R^2$ ) for triclosan and sparflaxacin were 0.9998 and 0.9999.

### 2.3. Distribution of drugs in the polymer

The distribution of the antimicrobials at the catheter surface after impregnation and after soaking in phosphate-buffered saline (PBS, Oxoid, Basingstoke, UK) for 12 weeks with weekly changes was studied by time of flight secondary ion mass spectrometry (ToF-SIMS) depth profiling. Transmission electron microscopy, on samples prepared by focussed ion beam milling or ultramicrotomy, was used to determine the distribution of the three drugs in the polymer matrix and to investigate their particle size. Elemental analysis was performed for the presence of nitrogen, fluorine and chlorine, expected to be present in the antimicrobials but absent from the polymer (Supplementary Method 2).

### Impregnation process



**Fig. 2.** Impregnation and elution of antimicrobials from catheter. Schematic depiction of the method used to produce antimicrobial catheters. Catheters were added to a solution of antimicrobials and given time to allow the solvent to swell the catheter (1 h). Catheters were then removed from solution and allowed to dry overnight, whereupon they returned to their original size.

#### 2.4. Mechanical properties of the catheter and balloon

It is important that any post-manufacture processing does not damage the biomaterial, and the inflatable retention balloon is an obvious site of susceptibility for any deterioration in mechanical properties. In addition, the catheter itself is subject to tensile force during insertion and removal. Urinary catheters were used in the “as-received” state from the manufacturers (controls) and following impregnation with the antimicrobial agents as described. Half of the control and antimicrobial catheters were placed in simulated urine [19] which was replaced weekly for up to 12 weeks to mimic drug elution in the body, while the other half remained in an unchanged state. Mechanical performance of catheter shafts was assessed using an Instron 5985 (Imetrum Ltd., Bristol, UK) with a 5 kN loadcell connected to a video extensometer. Data were recorded using Bluehill 2 software. A specialized clamping configuration was devised to connect the catheter tubing to the Instron as is shown in Fig. 6a. Catheters were clamped to expose a 5 cm length along the mid-section of the catheter and coated metal targets were applied 3 cm apart to allow for tracking by the video extensometer. A force was applied at a rate of 100 mm/min as the apparatus recorded the stress. The force applied was continued until either the apparatus reached its maximum length limit or until catheter failure (breakage) occurred. All tests were conducted at room temperature and in all cases three replicates of each analysis were performed. From the resultant relationship between the stress and strain, the load (N), ultimate tensile strength (MPa), elongation at break (mm) and modulus (MPa) determined at 280–320% elongation were calculated. Tests on catheter balloons (Supplementary Table 1) were adapted from ASTM F623 99 [20] and BSEN1616: 1997 [21]. The control and antimicrobial balloons were immersed in artificial urine [19] for 30 days at 37 °C. Balloon security was tested by inserting the catheter tips, with balloon inflated, through an inverted funnel and a 1 kg weight applied to pull the balloon against the neck of the funnel. Failure consists of leakage or rupture of the balloon and distortion so that the eye-holes of the catheter become occluded. All control and impregnated balloons, before and after soaking, passed the test. Balloon volume retention was tested by inflating them with 10 mL of 0.5 g/L methylene blue and placing on an absorbent white background (to detect dye leakage) for 30 days.

#### 2.5. Atomic force microscopy

An increase in surface roughness might facilitate mineral encrustation and could cause discomfort on catheter removal. Segments of control and impregnated catheters 1 cm long were placed in PBS at 37 °C for 12 weeks to simulate drug elution over the period of catheter use, with fresh solution replaced every week and a second identical set was left unsoaked. Atomic force microscopy was used to determine surface roughness (Rq value). All tests were performed in triplicate. Sections (approx 2 × 2 mm) of catheter segments were mounted onto metal stubs for analysis by AFM, and measurements were conducted on a MultiMode 8 AFM (Bruker) with a Nanoscope V controller operated in a

PeakForce Quantitative NanoMechanics (QNM) mode in air using a silicon nitride cantilever with a 0.1 N/m nominal spring constant. Two 5 μm × 5 μm surface area scans were performed on each catheter segment.

#### 2.6. Antimicrobial activity: serial plate transfer test

Test bacteria, chosen from isolates from patients with CAUTI in our urology clinic, were one strain each of: *Staphylococcus aureus* (MRSA), *E. coli*, *P. mirabilis*, *K. pneumoniae*, *Staphylococcus saprophyticus* and *Enterococcus faecalis*. In addition, we tested a strain of ESBL – producing *E. coli*, both *S. aureus* and *S. saprophyticus* were susceptible to rifampicin; *E. faecalis* had an MIC of rifampicin two logs higher, and the gram negative bacteria were predictably resistant to this drug. All were susceptible to triclosan except *E. faecalis*, which showed a low-level resistance. All were susceptible to sparflaxacin except *S. aureus* and the ESBL strain of *E. coli*. Minimum inhibitory concentrations (MICs) were determined by agar incorporation or in the case of rifampicin, by Etest (AB Biodisk, Solna, Sweden) and were shown in Supplementary Table 2. While a simple zone plate is sufficient to demonstrate antimicrobial activity, serial transfer of the material to fresh plates will show how long the material produces a zone of inhibition (Serial Plate Transfer Test, SPTT) [14]. ISA agar (Oxoid) plates were seeded with the test bacteria ( $A_{490}$  0.6,  $\sim 1 \times 10^7$  cfu/mL) and impregnated silicone discs were placed in triplicate on their surfaces and incubated overnight. Zones of inhibition were measured with calipers and the discs were transferred to a fresh seeded plate and incubated [22]. The process was repeated for 100 days or until no zones were produced (Fig. 5a). All tests were carried out in triplicate.

#### 2.7. Antimicrobial activity: time-kill study (tK100)

It is now recognized that, when bacteria attach to a surface, their susceptibility to antimicrobials is markedly reduced [23] and it is important to determine the ability of antimicrobial biomaterials to kill all, or most, of attached bacteria within a reasonable time. This assay has been termed the tK100 test, that is the time taken to kill 100% of attached bacteria. It is also recognized that when urinary catheters are contaminated by bacteria in vivo, the bacteria usually encounter a biomaterial surface that is overlaid with a host-derived conditioning film [24]. This was simulated by immersing the test discs and plain controls in filter-sterilized human urine (pH 6.8) for 1 h at 37 °C. XPS was used to determine if a conditioning film had been deposited on the surface (Supplementary Method 3). XPS confirmed the presence of a protein conditioning film (Fig. 5b). The discs were then immersed in a suspension (approximately  $1 \times 10^8$  colony-forming units/mL) of early log phase test bacteria and incubated at 37 °C for 1 h for attachment to take place. After rinsing to remove unattached bacteria, triplicates of discs were placed in diluted Tryptone Soya Broth (TSB, Oxoid) for up to 72 h, the dilution necessary for survival of attached controls without planktonic multiplication being found by experiment for each test isolate. For staphylococci 1% was suitable but for others 0.1%–0.5% was best (data not shown). At



intervals of 0, 24, 48 and 72 h, after rinsing and medium replacement each day, triplicates of discs were removed and sonicated (50 Hz for 20 min) and surviving colonies plate counted.

### 2.8. Antimicrobial activity: serial bacterial challenge in flow conditions


In order to simulate the flow aspects of the catheters in use and to ensure that any depletion of antimicrobial activity due to leaching was detected, we used an established catheter challenge model [18]. This consisted of a modular heating jacket into which the test catheters, with balloon removed, were inserted and kept hydrated by a secondary water jacket (Fig. 3a).

Tryptone Soya Broth was pumped through the catheters constantly at 0.5 mL/min (representative of human urine production rate) until the test ended. Each week, the test catheters (in triplicate) were inoculated with a suspension of  $\sim 1 \times 10^5$  cfu/mL of early log phase cultures of *S. aureus* (MRSA), *E. faecalis*, *E. coli* (including the ESBL strain) or *P. mirabilis* and the catheters clamped for 1 h to allow bacterial attachment (Fig. 3b). Fresh plain unimpregnated control catheters were included at each new inoculation. Flow was then restarted. Effluent samples from the distal end of the catheters were collected aseptically each day and plated (200  $\mu$ L per plate  $\times$  3) for colony counting. If no growth was present after 7 days, a further inoculation was made, and so on until either colonization occurred or 12 weekly challenges had passed. Colonized control catheters were removed and renewed for each challenge.

## 3. Results and discussion

### 3.1. Drug distribution and content of catheters and drug release profiles

The method of post-manufacture impregnation of Foley catheters described here differs from others in that it allows post-extrusion introduction of antimicrobials in a state that is not detectable by light or electron microscopy, and it allows the antimicrobials to migrate through the silicone matrix to replenish those lost from the catheter surface due to fluid flow, thus giving an extended duration of activity. Processed catheters were found to contain a total of 0.49 mg rifampicin (%RSD = 12.3), 14.0 mg triclosan (%RSD = 5.1) and 13.25 mg sparfloxacin (%RSD = 11.8) equivalent to 0.006%, 0.17% and 0.16% w/w respectively (Fig. 4a).



**Fig. 3.** Serial bacterial challenge in flow conditions. (a) Schematic depiction of the experimental setup for catheter challenge model experiments. (b) Example data of a standard experiment. Bacteria were injected weekly and bacterial numbers were measured daily at the sampling point. For each strain, measurements were taken for impregnated and untreated catheters.

The release profiles for the drugs from the silicone catheter segments over a 28 day period were determined (Fig. 4b). Release of rifampicin from the catheter segments was not detected (limit of detection 1  $\mu$ g/mL), probably because of the low initial total rifampicin content in the catheters. At time-points over the 28 day release, the %RSD of triclosan varied between 5.9–33.4% and 11.1–44.3% for sparfloxacin, with the higher RSD% being associated with concentrations of the drugs close to the analytical limits of detection. After 28 days in water, 30% of the initial triclosan and 20% of the initial sparfloxacin had been released, suggesting sufficient drug left to diffuse over a prolonged time.


In ToF-SIMS analysis, the Si<sup>-</sup> secondary ion was characteristic of the silicone substrate, the F<sup>-</sup> ion was characteristic for sparfloxacin and the Cl<sup>-</sup>/<sup>37</sup>Cl<sup>-</sup> ions were characteristic for triclosan (Supplementary Fig. 1). The CN<sup>-</sup> ion was representative for both sparfloxacin and rifampicin. Sparfloxacin and triclosan were shown to be evenly distributed laterally but before soaking there were some aggregates of both drugs at the surface, due to residues accrued during drying, that disappeared after soaking, though the drugs were still detected in the matrix (Fig. 4c). ToF-SIMS depth profiles of the impregnated catheters showed an elevated level of ions characteristic of drug compared with the untreated catheter (Fig. 4d), and a heightened intensity at the surface of the catheter. An even distribution of sparfloxacin and triclosan was observed throughout the remainder of the catheter wall (Fig. 4e). The presence of rifampicin could not be confirmed by ToF-SIMS for the catheter containing all three antimicrobials as representative ion for rifampicin (CN<sup>-</sup>) was also observed in the ToF-SIMS spectra for sparfloxacin. Therefore, to confirm that impregnation of rifampicin occurred, a catheter loaded only with this antimicrobial was assessed by ToF-SIMS depth profiling. An elevated intensity of the CN<sup>-</sup> ion was observed compared to an untreated catheter, which continued into the wall of the catheter, confirming that rifampicin was successfully loaded into the catheter using the methodology described. After a 12 week elution, a reduction in secondary ions characteristic of the drugs was observed throughout the catheter wall, suggesting that eluted drug is able to diffuse evenly from the catheter matrix (Fig. 4f) as intended.

TEM was unable to detect the antimicrobials on any scale (data not shown), and simulations indicated that the GIF Tridium TEM would be sensitive to localized drug concentrations of 10% or more by volume. This would indicate an upper size limit on aggregates of antimicrobials of 20 nm for the 200 nm-thick membrane analyzed in this study. This implies that the drugs are present with particle sizes of <20 nm and are probably in the form of a molecular dispersion rather than as granules as in the case when agents are added prior to extrusion.

### 3.2. Spectrum and duration of antimicrobial activity

The drug release profiles and antimicrobial activity show a long duration of activity, considerably in excess of other experimental antimicrobial catheters. Cho et al. dipped their catheters in a mixture of gentamicin and poly(ethylene-co-vinyl acetate but reported gentamicin release for only 7 days [25] and reported 5 days in a rabbit model [26]. The same group reported inhibition zones for 10 days but did not report a quantitative release profile [27]. Rafienia et al. reported 12 days' release from catheters dipped in a gentamicin-copolymer mix [28]. However, the extreme heterogeneity of antimicrobials, processing technology and test methods makes the published reports difficult to compare with each other and with our very different processing and testing. Our duration data are also far in excess of those reported for commercially available antimicrobial catheters [11,12] whose duration of activity is a few days at the most. Our previous studies on these two commercial catheters, silver-processed and nitrofur coated, showed a duration of activity of  $\leq 2$  days in the SPTT, and both failed the first challenge in the flow challenge test (data not shown), therefore validating the in vitro tests in respect of the clinical trial results.

In this SPTT study, discs impregnated with a combination of all three drugs (Fig. 5a) showed continued inhibition of all except *E. faecalis* for




**Fig. 4.** (a) Table of the amount of each drug loaded into the antimicrobial catheter, as determined by HPLC. (b) Drug elution over 28 day course for triclosan (■) and sparfloxacin (■), as determined by HPLC. (c) ToF-SIMS images of catheters untreated (left), after impregnation (middle) and after elution (right) for the total ion count,  $\text{Cl}^-$  characteristic of triclosan and  $\text{F}^-$  characteristic of sparfloxacin. The scale bar = 100  $\mu\text{m}$ . (d–f) ToF-SIMS depth profiles for (d) the untreated catheter, (e) impregnated catheter, and (f) catheter after elution. The area profiled was 100 Å–100  $\mu\text{m}$ . Secondary ions tracked were  $\text{Si}^-$  characteristic of silicone,  $\text{Cl}^-$  characteristic of triclosan and  $\text{F}^-$  characteristic of sparfloxacin.  $\text{CN}^-$  was also tracked, which is characteristic of both rifampicin and sparfloxacin. Ion intensity for  $\text{CN}^-$  from a catheter loaded with only rifampicin is shown in (e) as light green.

>100 days, with reductions of zone diameter of between 17% (staphylococci) and 36% (*P. mirabilis*) over this period. The inhibition zone for *E. faecalis* fell steeply and declined to zero on Day 22. No resistance was seen with any test bacterium in this assay. For the tK100 assay, we applied a urinary protein conditioning film. X-ray Photoelectron Spectroscopy (XPS), used to verify the deposition of a conditioning film, showed an increase in carbon and nitrogen with a decrease in silicon signal, indicating deposition of C + N rich material on the silicone surface (Fig. 5b). No viable *S. saprophyticus*, *S. aureus* or *K. pneumoniae* was found after 24 h. All *E. coli* were killed by 48 h and all *E. faecalis* by 72 h. In the case of *P. mirabilis*, <25 cfu/mL remained at 72 h compared to  $10^5$  cfu/mL from plain controls indicating a >99.9% reduction (Fig. 5c and d). In the flow challenge test, on each challenge, all the control catheters became colonized by Day 1 and remained so until removed (Table 1).

No colonies were observed for *S. aureus* and *E. coli*, including the ESBL strain in effluents for the 84-day duration of the challenge test. Those challenged with *P. mirabilis* became colonized on Days 57, 78 and 85. The processed catheters failed to prevent colonization by *E. faecalis* beyond Day 15. Repeat MIC determination on isolates from failed catheters showed slight increases for *P. mirabilis* but still within the susceptible range, and no change for *E. faecalis*, suggesting that in each case failure to prevent colonization was not due to the development of resistance. Resistance development was not seen after prolonged exposure of test bacteria to the catheters in the Serial Plate Transfer Test or in the Serial Challenge Flow Test. Concerns regarding

resistance strains appearing as a result of use of antimicrobial catheters are entirely justified and this is why we used these combinations of antimicrobials, which according to the Dual Drug Principle [29] would significantly reduce the likelihood of mutational resistance. Also, when much higher inocula of *P. mirabilis* ( $\sim 10^8$  cfu/mL) have been exposed to triclosan alone on agar surfaces they have shown evidence of mutational resistance, as expected [30]. However, those mutants showing MICs of 2 mg/L were still prevented from colonizing urinary catheters in a laboratory model [30]. In most cases reduction of viable bacteria was below the detectable levels, at least a five-log reduction (Table 1: in vitro challenge, and Fig. 5c and d, tK100). Overall, the test catheters were able to delay colonization by CAUTI pathogens between 7 and 12 weeks, with the exception of *E. faecalis*. The spectrum of activity therefore extends to most of the CAUTI uropathogens, the exceptions being *P. aeruginosa*, *Candida* and, except for the first two weeks, enterococci. However, most CAUTI is caused by enterobacteria (*E. coli*, *K. pneumoniae*, *P. mirabilis*) with a minority due to *P. aeruginosa* or *Candida* in most studies [31].

Long-term urinary catheters have also been identified as reservoirs for *S. aureus* including MRSA [32] and our catheter also showed a strong activity against these bacteria. As well as reducing CAUTI due to staphylococci, the catheter promises to reduce spread of *S. aureus* and MRSA, contributing to infection control measures. The results suggest that this might similarly be the case for multi-resistant ESBL *E. coli*, reducing the spread of these bacteria to the environment, to healthcare workers and to other patients.



**Fig. 5.** Biological characterization of impregnated catheters. (a) Serial zone of inhibition analysis (Serial Plate Transfer Test) for impregnated silicone discs on an agar surface for *E. faecalis* ( $\blacktriangle$ ), *K. pneumoniae* ( $\ast$ ), *P. mirabilis* ( $\bullet$ ), *S. aureus* ( $\blacklozenge$ ), *E. coli* ( $\times$ ), *S. saprophyticus* ( $\blacksquare$ ), and ESBL *E. coli* ( $\blackplus$ ). (b) Atomic % of impregnated and untreated catheters with and without conditioned films, as determined by XPS. (c–d) tK100 assay on impregnated (solid line) and untreated (dotted line) catheters for (c) *E. faecalis* ( $\blacktriangle$ ), *S. aureus* ( $\blacklozenge$ ), *S. saprophyticus* ( $\blacksquare$ ), and (d) *K. pneumoniae* ( $\ast$ ), *P. mirabilis* ( $\bullet$ ), and *E. coli* ( $\times$ ).

### 3.3. Effect of processing on mechanical properties of the catheter and balloon

The results show that neither the process itself nor the presence of the antimicrobials affects the mechanical properties of the material. Impregnating the catheter and balloon with antimicrobial agents caused no adverse effects to the mechanical performance (Fig. 6b, c).

Despite a load of 60–70 N applied at 100 mm/min, resulting in an elongation of 377% (instrument limit) for all catheters, none was fractured. There were no statistically significant differences (2 tailed homoscedastic Student's t-test) between control and impregnated catheters in their unused state with regard to load ( $p = 0.168$ ), tensile strength ( $p = 0.111$ ), and modulus ( $p = 0.054$ ) or between control and impregnated catheters after soaking in artificial urine for load ( $p = 0.993$ ), tensile strength ( $p = 0.994$ ) and modulus ( $p = 0.381$ ) (Fig. 6c). All control and impregnated balloons, before and after soaking, passed the test. No

leakage was seen from control or impregnated balloons, and in all cases the balloons deflated fully at the end of the test (Supplementary Table 1), an essential requirement in clinical use to allow free removal of the catheter from the bladder. Impregnating the catheter and balloon with antimicrobial agents therefore caused no adverse effects to the mechanical performance. The processing and impregnation did not change the surface roughness of the catheters. There was no significant difference (2 tailed homoscedastic Student's t-test) in Rq values between control and impregnated catheters before soaking ( $p = 0.240$ ), but after soaking the impregnated catheters showed an increase in the Rq value ( $9.57 \text{ nm} \pm 1.35 \text{ s.d.}$  vs  $22.72 \text{ nm} \pm 4.33 \text{ s.d.}$ ,  $p = 1.628 \times 10^{-15}$ ), presumably due to the release of antimicrobials during soaking (Fig. 6d). This nm increase in surface roughness would be unlikely to cause patient discomfort, or act to promote bacterial colonization or mineral encrustation, though further testing of mineral encrustation propensity would be required to exclude this possibility.



**Table 1**

Results of the serial flow catheter challenge. During constant flow, catheters (in triplicate) were challenged weekly with a suspension of  $\sim 1 \times 10^5$  cfu/mL of early log phase cultures of *S. aureus* (MRSA), *E. faecalis*, *E. coli* (including the ESBL strain) or *P. mirabilis* and the catheters clamped for 1 h to allow bacterial attachment. After resumption of flow, effluents were plated daily for quantitation. Challenges were repeated until colonization occurred or until 12 successive challenges had been administered. Control (plain) catheters were renewed at each challenge. All control catheters showed colonization within 24 h. Catheters withstood challenge with *S. aureus*, *E. coli* and ESBL *E. coli* for at least 12 challenges (84 days), while the three challenged with *P. mirabilis* became colonized after 57, 78 and 84 days respectively. Protection against colonization by *E. faecalis* was shortlived, lasting for 15 days.

Test bacterium	Day of failure
<i>Staphylococcus aureus</i> MRSA	>84
<i>Enterococcus faecalis</i>	15
<i>Escherichia coli</i>	>84
<i>Escherichia coli</i> ESBL	>84
<i>Proteus mirabilis</i>	57, 78, 84

#### 4. Conclusions

Currently available “antimicrobial” urinary catheters have been shown not to be effective in laboratory and clinical studies, and there is no catheter available for either short-term or long-term catheter users, leaving them with the problem of recurring infection and obstruction leading to painful catheter changes. If the antimicrobial

catheter were to prove to be effective in clinical use then a considerable saving in hospital time and treatment and vastly improved quality of life for patients could be achieved. The results suggest that the impregnated catheters might reduce CAUTI in both short-term and long-term urinary catheter use. Though this reduction is difficult to quantify at this stage, it is likely that, if CAUTI could be prevented or significantly delayed in a new catheter user, the concomitant reduction in antibiotic use might also avoid progression to infection by *Pseudomonas* or *Candida*, and in any case would contribute considerably to the drive to reduce antibiotic resistance.


Further studies on the antimicrobial catheter will include an evaluation of its ability to reduce or prevent mineral encrustation, before a Phase 1 study followed by a full Phase 2 clinical efficacy trial.

#### Contributions

RB and LEF conceived and designed the experiments. LEF, ALH, WA, RB, AY, DAB, DJS, XC, MF, EFS and CDJP carried out the experiments and supplied specialist expertise. RP gave specialist urological advice. LEF, RB and ALH wrote the paper and all authors reviewed the draft.

#### Competing financial interests

RB has filed a patent application, approved in several countries, on the impregnation technology. No other author has any competing financial interest.



**Fig. 6.** Mechanical characterization of impregnated catheters. (a) Image of the setup used to assess the load, tensile strength and modulus of catheters. (b) Image of impregnated (left) and untreated (right) catheters with the balloon inflated. (c) Table of the mechanical measurements taken for untreated and impregnated catheters before and after elution. (d) AFM characterization of untreated and impregnated catheters before (left) and after (right) elution, showing both the topographical images and RMS (Root Mean Square) roughness ( $R_q$ ). Images were acquired from an area of  $5 \mu\text{m} \times 5 \mu\text{m}$ .

## Acknowledgments

Paul Cooling is gratefully acknowledged for HPLC assistance, Keith Dinsdale and Tom Buss for mechanical property testing guidance and Dr Ignacio Villar for XPS assistance. LEF received a studentship from the University of Nottingham.

## Appendix A. Supplementary data

Supplementary data to this article can be found online at <http://dx.doi.org/10.1016/j.jconrel.2015.01.037>.

## References

- [1] K.N. Moore, K.F. Hunter, R. McGinnis, C. Bacsu, M. Fader, et al., Do catheter washouts extend patency time in long-term indwelling urethral catheters? *J. Wound Ostomy Continence Nurs.* 36 (2009) 82–90.
- [2] D. Cardo, T. Horan, M. Andrus, M. Dembinski, J. Edwards, G. Peavy, et al., National Nosocomial Infections Surveillance (NNIS) System Report, data summary from January 1992 through June 2004, issued October 2004, *Am. J. Infect. Control* 32 (2004) 470–485.
- [3] M.A. Dudeck, L.M. Weiner, K. Allen-Bridson, P.J. Malpiedi, K.D. Peterson, D.A. Pollock, et al., National Healthcare Safety Network (NHSN) report, data summary for 2012, device-associated module, *Am. J. Infect. Control* 41 (2013) 1148–1166.
- [4] D.J. Stickler, S.D. Morgan, Observations on the development of the crystalline bacterial biofilms that encrust and block Foley catheters, *J. Hosp. Infect.* 69 (2008) 350–360.
- [5] S.M. Jacobsen, D.J. Stickler, H.L.T. Mobley, M.E. Shirtliff, Complicated catheter-associated urinary tract infections due to *Escherichia coli* and *Proteus mirabilis*, *Clin. Microbiol. Rev.* 21 (2008) 26–59.
- [6] P.F. Holmes, E.P. Currie, J.C. Thies, H.C. van der Mei, H.J. Busscher, W. Norde, Surface-modified nanoparticles as a new, versatile, and mechanically robust nonadhesive coating: suppression of protein adsorption and bacterial adhesion, *J. Biomed. Mater. Res.* 91A (2009) 824–833.
- [7] A. Roosjen, H.J. Kaper, H.C. van der Mei, W. Norde, H.J. Busscher, Inhibition of adhesion of yeasts and bacteria by poly(ethylene oxide)-brushes on glass in a parallel plate flow chamber, *Microbiology* 149 (2003) 3239–3246.
- [8] A.L. Hook, C.Y. Chang, J. Yang, J. Luckett, A. Cockayne, S. Atkinson, et al., Combinatorial discovery of polymers resistant to bacterial attachment, *Nat. Biotechnol.* 30 (2012) 868–875.
- [9] A.L. Hook, C.Y. Chang, J. Yang, S. Atkinson, R. Langer, D.G. Anderson, et al., Discovery of novel materials with broad resistance to bacterial attachment using combinatorial polymer microarrays, *Adv. Mater.* 25 (2013) 2542–2547.
- [10] J. Davenas, P. Thevenard, F. Philippe, M.N. Arnaud, Surface implantation treatments to prevent infection complications in short term devices, *Biomol. Eng.* 19 (2002) 263–268.
- [11] J.H. Crabtree, R.J. Burchette, R.A. Siddiqi, I.T. Huen, L.L. Hadnott, A. Fishman, Efficacy of silver-ion implanted catheters in reducing peritoneal dialysis-related infections, *Perit. Dial. Int.* 23 (2003) 368–374.
- [12] R. Pickard, T. Lam, G. MacLennan, K. Starr, M. Kilonzo, G. McPherson, et al., Antimicrobial catheters for reduction of symptomatic urinary tract infection in adults requiring short-term catheterisation in hospital: a multicentre randomised controlled trial, *Lancet* 380 (2012) 1927–1935.
- [13] T.M. Hooton, S.F. Bradley, D.D. Cardenas, R. Colgan, S. Geerlings, J.C. Rice, et al., Diagnosis, prevention, and treatment of catheter-associated urinary tract infection in adults: 2009 International Clinical Practice Guidelines from the Infectious Diseases Society of America, *50* (2010) 625–663.
- [14] R. Bayston, N. Grove, J. Siegel, D. Lawellin, S. Barsham, Prevention of hydrocephalus shunt catheter colonization in vitro by impregnation with antimicrobials, *J. Neurosurg. Neurosurg. Psychiatry* 52 (1989) 605–609.
- [15] S.T. Govender, N. Nathoo, J.R. van Dellen, Evaluation of an antibiotic-impregnated shunt system for the treatment of hydrocephalus, *J. Neurosurg.* 99 (2003) 831–839.
- [16] H.E. Aryan, H.S. Meltzer, M.S. Park, R.L. Bennett, R. Jandial, M.L. Levy, Initial experience with antibiotic-impregnated silicone catheters for shunting of cerebrospinal fluid in children, *Childs Nerv. Syst.* 21 (2005) 56–61.
- [17] D.M. Sciubba, R.M. Stuart, M.J. McGirt, G.F. Woodworth, A. Samdani, B. Carson, et al., Effect of antibiotic-impregnated shunt catheters in decreasing the incidence of shunt infection in the treatment of hydrocephalus, *J. Neurosurg.* 103 (2005) 131–136.
- [18] R. Bayston, L.E. Fisher, K. Weber, An antimicrobial modified silicone peritoneal catheter with activity against both Gram positive and Gram negative bacteria, *Biomaterials* 30 (2009) 3167–3173.
- [19] D.P. Griffith, D.M. Musher, C. Itin, Urease – primary cause of infection-induced urinary stones, *Investig. Urol.* 13 (1976) 346–350.
- [20] ASTM, F623-99 (Reapproved 2006), Standard Performance Specification for Foley Catheter, ASTM International, 2006.
- [21] S.H. Yuk, S.H. Cho, S.H. Lee, pH/temperature-responsive polymer composed of poly((N,N-dimethylamino)ethyl methacrylate-co-ethylacrylamide), *Macromolecules* 30 (1997) 6856–6859.
- [22] R. Bayston, R.D.G. Milner, Anti-microbial activity of silicone-rubber used in hydrocephalus shunts, after impregnation with anti-microbial substances, *J. Clin. Pathol.* 34 (1981) 1057–1062.
- [23] W.L. Cochran, G.A. McFeters, P.S. Stewart, Reduced susceptibility of thin *Pseudomonas aeruginosa* biofilms to hydrogen peroxide and monochloramine, *J. Appl. Microbiol.* 88 (2000) 22–30.
- [24] D.J. Denstedt, T.A. Wollin, G. Reid, Biomaterials used in urology: current issues of biocompatibility, infection, and encrustation, *J. Endourol.* 12 (1998) 493–500.
- [25] Y.W. Cho, J.H. Park, S.H. Kim, Y.H. Cho, J.M. Choi, H.J. Shin, et al., Gentamicin-releasing urethral catheter for short-term catheterization, *J. Biomater. Sci. Polym. Ed.* 14 (2003) 963–972.
- [26] Y.-H. Cho, S.-J. Lee, J.Y. Lee, S.W. Kim, I.C. Kwon, S.Y. Chung, M.S. Yoon, Prophylactic efficacy of a new gentamicin-releasing urethral catheter in short-term catheterized rabbits, *BJU Int.* 87 (2001) 104–109.
- [27] J.H. Park, Y.W. Cho, Y.H. Cho, J.M. Choi, H.J. Shin, Y.H. Bae, et al., Norfloxacin-releasing urethral catheter for long-term catheterization, *J. Biomater. Sci. Polym. Ed.* 14 (2003) 951–962.
- [28] M. Rafienia, B. Zarinmehr, S.A. Poursamar, S. Bonakdar, M. Ghavami, M. Janmaleki, Coated urinary catheter by PEG/PVA/gentamicin with drug delivery capability against hospital infection, *Iran. Polym. J.* 22 (2012) 75–83.
- [29] X. Zhao, K. Drlica, Restricting the selection of antibiotic-resistant mutants: a general strategy derived from fluoroquinolone studies, *Clin. Infect. Dis.* 33 (Suppl. 3) (2001) S147–S156.
- [30] D.J. Stickler, G.L. Jones, Reduced susceptibility of *Proteus mirabilis* to triclosan, *Antimicrob. Agents Chemother.* 52 (2008) 991–994.
- [31] L.E. Nicolle, The chronic indwelling catheter and urinary tract infection in long-term-care facility residents, *Infect. Control Hosp. Epidemiol.* 22 (2001) 316–321.
- [32] R.R. Muder, C. Brennan, J.D. Rihs, M.M. Wagener, A. Obman, J. Stout, et al., Isolation of *Staphylococcus aureus* from the urinary tract: association of isolation with symptomatic urinary tract infection and subsequent staphylococcal bacteremia, *Clin. Infect. Dis.* 42 (2006) 46–50.